




XJ0000178



**ОБЪЕДИНЕННЫЙ  
ИНСТИТУТ  
ЯДЕРНЫХ  
ИССЛЕДОВАНИЙ**

**Дубна**

E16-2000-144

V.E.Aleinikov\*, L.G.Beskrovnaja, B.V.Florko

**CHARACTERISTICS  
OF POLYETHYLENE-MODERATED  
<sup>252</sup>Cf NEUTRON SOURCES**

Submitted to the International Workshop  
on «Neutron Field Spectrometry in Science, Technology  
and Radiation Protection», June 4–8, 2000, Piza, Italy

\*Corresponding author. Fax: (007 09621) 65 166;  
E-mail: aleinik@jinr.ru

**31 / 45**

**2000**

## **1. Introduction**

It is widely recognised that neutron dosimeters do not exhibit a constant response factor (per unit dose equivalent) as neutron energies and spectral distributions are varied from one working environment to another. In addition, most dosimeters cannot distinguish one neutron spectrum from another. Algorithms based on the response of a dosimeter cannot, therefore, be developed to determine the appropriate response factor to be applied to a neutron dosimeter irradiated by an unknown neutron spectrum. Thus, it is necessary to determine a response factor for each type of neutron spectrum in which each type of neutron dosimeter will be used.

An alternative (and somewhat more practical) approach is to design a set of reference neutron fields (RF) that simulate the neutron spectra to which a generic group of radiation personnel might be exposed. This set of RF should cover, in practice, the entire interval of neutron energies generated by nuclear installations.

To this end, a set of RF has been created in the Joint Institute for Nuclear Research (JINR) that would produce spectra similar to the neutron spectra observed in the workplaces around high energy accelerators and nuclear reactors. The set comprises reference fields based on a  $^{252}\text{Cf}$  source at the centre of polyethylene moderators 12.7 and 29.2 cm in diameter as well as two reference fields based on a 600 MeV phasotron[1].

The paper reports the calculated neutron spectra and dosimetric characteristics of the RF based on polyethylene - moderated  $^{252}\text{Cf}$  neutron sources.

## 2. Reference fields

Reference fields were created using a  $^{252}\text{Cf}$  source placed at the centre of polyethylene spheres with diameters 12.7 and 29.2 cm. The fields are produced inside a room  $5.2 \times 10.8 \times 3.5 \text{ m}^3$  and take into account the scattered neutrons. The moderators with the source are placed on a calibration stand and the fields are established at a distance of 100 cm from the source centre. In this case it is not necessary to provide special monitoring. There is a need only to maintain good background conditions in the room and to take into account the source decay.

## 3. Neutron spectra and conversion coefficients

Neutron spectra were calculated by Monte Carlo method[2] and compared with experimental data [1]. The calculated spectra of the fields free in air and in the calibration room from  $^{252}\text{Cf}$  source (with yield  $1 \text{ n}\cdot\text{s}^{-1}$ ) placed at the centre of polyethylene spheres with diameters 12.7cm and 29.2 cm as well as unfolded spectra[1] from count rates of Bonner sphere detectors are shown in figures 1 and 2. Conversion coefficients for monoenergetic neutrons that relate neutron fluence to the operational quantities ambient dose equivalent and personal dose equivalent have been summarized in ICRU Report 57[3]. The fluence-to-dose equivalent coefficient,  $h^*\phi$ , for a neutron spectrum can be evaluated using the following formula:

Tables 1.

Ambient and personal dose equivalent per unit neutron fluence, averaged over neutron spectra of CH<sub>2</sub>-moderated <sup>252</sup>Cf (diameter of moderator is 12.7 cm).

Neutron source position	$h^*_{\phi}(10)$	$h_{p,\phi}(10,\alpha)$ $\alpha = 0^{\circ}$	$h_{p,\phi}(10,\alpha)$ $\alpha = 15^{\circ}$	$h_{p,\phi}(10,\alpha)$ $\alpha = 30^{\circ}$	$h_{p,\phi}(10,\alpha)$ $\alpha = 45^{\circ}$	$h_{p,\phi}(10,\alpha)$ $\alpha = 60^{\circ}$	$h_{p,\phi}(10,\alpha)$ $\alpha = 75^{\circ}$
Free in air	176	188	186	192	179	158	105
In calibration room	120	128	126	130	120	104	68

Tables 2.

Ambient and personal dose equivalent per unit neutron fluence, averaged over neutron spectra of CH<sub>2</sub>-moderated <sup>252</sup>Cf (diameter of moderator is 29.2 cm).

Neutron source position	$h^*_{\phi}(10)$	$h_{p,\phi}(10,\alpha)$ $\alpha = 0^{\circ}$	$h_{p,\phi}(10,\alpha)$ $\alpha = 15^{\circ}$	$h_{p,\phi}(10,\alpha)$ $\alpha = 30^{\circ}$	$h_{p,\phi}(10,\alpha)$ $\alpha = 45^{\circ}$	$h_{p,\phi}(10,\alpha)$ $\alpha = 60^{\circ}$	$h_{p,\phi}(10,\alpha)$ $\alpha = 75^{\circ}$
Free in air	127	135	134	138	129	115	79
In calibration room	90	96	95	97	90	78	52




Fig 1. The neutron spectra of the fields free in air and in the calibration room from  $^{252}\text{Cf}$  source (with yield  $1 \text{ n}\cdot\text{s}^{-1}$ ) placed at the centre of polyethylene spheres with diameter 12.7 cm.

1 - calculated spectrum in the calibration room; 2 - unfolded spectrum in the calibration room;

3 - calculated spectrum free in air.




Fig 2. The neutron spectra of the fields free in air and in the calibration room from  $^{252}\text{Cf}$  source (with yield  $1 \text{ n}\cdot\text{s}^{-1}$ ) placed at the centre of polyethylene spheres with diameter 29.2 cm.

1 - calculated spectrum in the calibration room; 2 - unfolded spectrum in the calibration room;

3 - calculated spectrum free in air.

$$h_{\Phi}^* = \frac{\int h(E)\Phi(E)dE}{\int \Phi(E)dE} \quad , \quad (1)$$

where  $h(E)$  is conversion coefficients for monoenergetic neutrons,  $\text{pSv}\cdot\text{cm}^2$ ;  $\Phi(E)$  is neutron spectrum and  $E$  is neutron energy.

Table 1 and 2 show ambient and personal dose equivalent per unit neutron fluence,  $h_{\Phi}^*(10) = H^*(10)/\Phi$  and  $h_{p,\Phi}(10,\alpha) = H_{p,\text{slab}}(10,\alpha)/\Phi$  ( $\text{pSv}\cdot\text{cm}^2$ ), averaged over the neutron spectra.

#### 4. Concluding remarks

The neutron reference fields produce neutron spectra that are similar to the spectra observed in the workplace of a major category of workers. When used properly, these fields provide a convenient method to calibrate neutron dosimeters and to conduct neutron dosimetry intercomparison studies.

#### References

- [1] V.E. Aleinikov et al., *Radiat. Prot. Dosim.* 54 (1994) 57.
- [2] MCNP - A General Monte Carlo N-Particle Transport Code, La-12625-M, Version 4B, (1997).
- [3] ICRU Report 57, International Commission on Radiation Units and Measurements, Bethesda, Maryland, (1998).

Received by Publishing Department  
on June 20, 2000.

**The Publishing Department  
of the Joint Institute for Nuclear Research  
offers you to acquire the following books:**

Index	Title
E1,2-97-79	Proceedings of the XIII International Seminar on High Energy Physics Problems. Relativistic Nuclear Physics and Quantum Chromodynamics. Dubna, 1996 (2 volumes, 364 p. and 370 p., in English)
D5,11-97-112	Proceedings of the 9th International Conference «Computational Modelling and Computing in Physics». Dubna, 1996 (378 p., in English)
E2-97-213	Proceedings of the V International Seminar on Interaction of Neutron with Nuclei «Neutron Spectroscopy, Nuclear Structure, Related Topics». Dubna, 1997 (446 p., in English)
E2,4-97-263	Proceedings of the Third International Conference «Renormalization Group'96». Dubna, 1996 (436 p., in English)
E10-97-272	Proceedings of the Data Acquisition Systems of Neutron Experimental Facilities (DANEF'97). Dubna, 1997 (325 p., in English)
D19-97-284	Proceedings of the International Symposium «Problems of Biochemistry. Radiation and Space Biology». Dubna, 1997 (2 volumes 284 p. and 405 p., in Russian and English)
E2-97-413	Proceedings of the VII Workshop on High Energy Spin Physics (SPIN'97). Dubna, 1997 (398 p., in English)
E15-98-57	Proceedings of the III International Workshop «Hyperfine Structure and Nuclear Moments of Exotic Nuclei by Laser Spectroscopy». Poznan, 1997 (200 p., in English)
D13-98-66	Proceedings of the XVII International Symposium on Nuclear Electronics. Varna, 1997 (242 p., in English and Russian)
E1,2-98-154	Proceedings of the XIII International Seminar on High Energy Physics Problems «Relativistic Nuclear Physics and Quantum Chromodynamics». Dubna, 1996 (2 volumes 300 p. and 282 p., in English)
E3,14-98-168	Proceedings of the German–Russian User Meeting «Condensed Matter Physics with Neutrons at IBR-2». Dubna, 1998 (126 p., in English)
E3-98-202	Proceedings of the VI International Seminar on Interaction of Neutrons with Nuclei Neutron Spectroscopy, Nuclear Structure, Related Topics. Dubna, 1998 (352 p., in English)
D1,2-98-215	Proceedings of the International Workshop «Relativistic Nuclear Physics: from MeV to TeV». Dubna, 1998 (384 p., in English and Russian)
E2-98-254	Proceedings of the International Workshop Hadronic Atoms and Positronium in the Standard Model. Dubna, 1998 (260 p., in English)



Index	Title
D9,11-98-273	Proceedings of the 4th International Workshop «Beam Dynamics and Optimization». Dubna, 1997 (162 p., in English and Russian)
E17,19-98-305	Proceedings of the International Workshop on Deuteration of Biological Molecules for Structural and Dynamic Studies. Applications to Neutron Scattering and NMR. Dubna, 1998 (100 p., in English)
E1,2-98-307	Proceedings of the International School–Seminar «Actual Problems of Particle Physics». Gomel, 1997, Belarus (2 volumes 304 p. and 220 p., in English)
E2-98-372	Proceedings of the III International Workshop «Classical and Quantum Integrable Systems». Erevan, 1998 (200 p., in English)
E9-99-26	Proceedings of the International Conference on High Energy Accelerators. Dubna, 1998 (432 p., in English)
E2-99-35	Proceedings of the XI International Conference on Problems of Quantum Field Theory. Dubna, 1998 (508 p., in English)
E5-99-38	Proceedings of the International Workshop «Self-Similar Systems». Dubna, 1998 (404 p., in English)
E9-99-92	Proceedings of the XVII International Conference on High Energy Accelerators HEACC'98. Dubna, 1998 (435 p., in English)

Please apply to the Publishing Department of the Joint Institute for Nuclear Research for extra information. Our address is:

Publishing Department  
 Joint Institute for Nuclear Research  
 Dubna, Moscow Region  
 141980 Russia  
 E-mail: [publish@pds.jinr.dubna.su](mailto:publish@pds.jinr.dubna.su).

Алейников В.Е., Бескровная Л.Г., Флорко Б.В.  
Характеристики нейтронных источников на основе  $^{252}\text{Cf}$   
в полиэтиленовых замедлителях

E16-2000-144

Нейтронные источники на основе  $^{252}\text{Cf}$  в полиэтиленовых замедлителях были созданы для получения спектров нейтронов в опорных полях, близких по своим характеристикам к спектрам, имеющимся на ядерных реакторах и ускорителях. Описываются источники нейтронов и их поля. Нейтронные спектры были рассчитаны методом Монте-Карло и сравнивались с экспериментальными данными.

Работа выполнена в Отделении радиационных и радиобиологических исследований ОИЯИ.

Препринт Объединенного института ядерных исследований. Дубна, 2000

Aleinikov V.E., Beskrovnaja L.G., Florco B.V.  
Characteristics of Polyethylene-Moderated  $^{252}\text{Cf}$  Neutron Sources

E16-2000-144

Polyethylene-moderated  $^{252}\text{Cf}$  neutron sources were designed to produce neutron reference fields' spectra that simulate the spectra observed in the workplaces within nuclear reactors and accelerators. The paper describes the neutron sources and fields. Neutron spectra were calculated by Monte Carlo method and compared with experimental data.

The investigation has been performed at the Division of Radiation and Radiobiological Research, JINR.

Preprint of the Joint Institute for Nuclear Research. Dubna, 2000

Макет Т.Е.Попеко

Подписано в печать 30.06.2000  
Формат 60 × 90/16. Офсетная печать. Уч.-изд. листов 1,04  
Тираж 240. Заказ 52116. Цена 1 р. 25 к.

Издательский отдел Объединенного института ядерных исследований  
Дубна Московской области